



Owlstone Permeation Source Certificate of Calibration

Material: Formaldehyde
Part No.: 70-0027

Source Serial No.: 100-499

Type: Tubular

Vapour: Formaldehyde

Calibration Temperature: 50°C

Permeation Rate: 1520 ng min⁻¹

Date of Calibration: 01st June 2012

Calibration Method: Gravimetric

Russell Parris
Chief Applications Officer

Dated